



Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
Ecolab®	A-456-II		Minimum 256 to 1 dilution with water	02-Dec-14
Quantum Pacific Limited	Actimax		According to Manufacturer Spec	11-Apr-06
SCIENCECompany	Ammonium Chloride	Cocoalkaydimethylbenzyl ammonium chloride	According to Manufacturer Spec	17-May-19
Milton®	Anti-Bacterial Tablets	Sodium Dichloroisochyanurate 30-60%	1 Tablet:2L Water	19-Jul-12
Ecolab®	Asepti Active Liquid		According to Manufacturer Spec	20-Nov-12
Ecolab®	Asepti Active Wipes	Dual-Quaternary/Alcohol	Ready to Use	20-Nov-12
Synergy Health	Azomax™ Multi Surface Wipes	Benzylkoniumchloride 21.6%	Ready to Use	15-Mar-13
Synergy Health	Azowipe™ Hard Surface Wipes	Benzylkoniumchloride 2%	Ready to Use	15-Mar-13
PreVasive®	BAC Botanical Antimacrobial Cleaner	Thyme Oil	According to Manufacturer Spec	22-May-19
Spartan Chemical Company Inc.	BNC-15™		According to Manufacturer Spec	11-Aug-16
Metrex® Research	CaviCide [®]	Isopropanol	According to Manufacturer Spec	06-Oct-16
Metrex® Research	CaviCide1™	Isopropanol	According to Manufacturer Spec	05-Jan-17
Metrex® Research	CaviWipes1™	Isopropanol	Ready to Use	05-Jan-17
Metrex® Research	CaviWipes™	Isopropanol	Ready to Use	06-Oct-16
Agar Cleaning Systems Pty. Ltd.	Chloradet	Sodium Hypochlorite <10%	Minimum 1000 to 1 dilution with water	06-Jun-16
Apollo Corporation	CID A-L™ II Quaternary Disinfectant	Quaternary Ammonium Compounds	According to Manufacturer Spec	26-Sep-12

[%] Tel: 64 9 274 0869

^{⊚ 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

^{⊕ 1} Spireton Place, Pendle Hill, NSW 2145, Australia

PO Box 335, Pendle Hill, NSW 2145, Australia





Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
Wexford Labs, Inc.	CleanCide® Wipes	Citric Acid	Ready to Use	15-May-19
Clorox Disinfecting Wipes	Clorox Disinfecting Wipes		According to Manufacturer Spec	14-Feb-11
Alco Mandala Products	DC-7		According to Manufacturer Spec	16-Feb-12
Clinell®	Detergent Wipes	Propane 1,2-diol	According to Manufacturer Spec	27-May-09
ArjoHuntleigh	Disinfectant Cleanser IV		According to Manufacturer Spec	23-Nov-09
Clorox®	Dispatch® Disinfectant		According to Manufacturer Spec	16-Feb-12
Tristel Solutions LTD	Distel Surface Disinfectant (Formerly Trigene®)	Polymeric Biguanide Hydrochloride <10%	According to Manufacturer Spec	24-Oct-06
Sealed Air Diversey Care	Divercleanse	Sodium Hypochlorite <5%, Sodium Peroxide <2%	Minimum 33 to 1 dilution with water	15-Dec-09
Buckeye® International, Inc.	Eco E23 S23	5.6	Minimum 256 to 1 dilution with water	05-Jan-17
Velocity Chemicals	Eco-Pine		Minimum 20 to 1 dilution with water	11-Apr-06
Clorox HEALTHCARE®	Fuzion™		According to Manufacturer Spec	05-Jan-17
Spartan Chemical Company Inc.	GS Neutral Disinfectant #3502		According to Manufacturer Spec	02-Dec-14
Clinell®	Hand & Surface Sanitizer	Benzalkonium Chloride	According to Manufacturer Spec	27-May-09
3M	HB Quat Disinfectant	Alkyl dimethyl benzyl ammonium chloride	Ready to Use	06-Sep-19
Spartan Chemical Company Inc.	hdq®C2	Ammonium Chloride	According to Manufacturer Spec	30-Jun-17
Wipesplus	Hospital/Surgical Wipes	Ammonium Chloride	Ready to Use	15-May-19

[∜] Tel: 64 9 274 0869

^{⊚ 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

^{⊗ 1} Spireton Place, Pendle Hill, NSW 2145, Australia

PO Box 335, Pendle Hill, NSW 2145, Australia





Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
Clorox HEALTHCARE®	Hydrogen Peroxide Cleaner Disinfectant Wipes	Hydrogen Peroxide	Ready to Use	12-Apr-19
Apollo Corporation	Hygena™	Ethylalcohol 1.4%	According to Manufacturer Spec	25-Sep-12
Kimberly-Clark®	Isowipes®	Isopropanol 30-60%	Ready to Use	19-Jul-12
Legge Systems	Legcide HB Plus		According to Manufacturer Spec	*****
Whiteley Medical	Lemex		According to Manufacturer Spec	15-Mar-11
Livingstone International (PTY) LTD	LIV-Wipes Jumbo	70% Isopropyl Alcohol	Ready to Use	15-Mar-11
Sulco LTD	Mediwipes		Ready to Use	27-May-09
Medline Industries Inc.	MICRO-KILL™ + Disinfecting, Deodorizing, Cleaning Wipes	Isopropyl Alchohol	According to Manufacturer Spec	09-Mar-20
Medline Industries Inc.	MICRO-KILL™ AF2 Wipes	Ammonium Chloride	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ Bleach-Free, Alcohol- Free Wipes	Ammonium Chloride	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ One Wipes	Ethyl Alchohol	Ready to Use	15-May-19
Medline Industries Inc.	MICRO-KILL™ Q3 Concentrated Disinfectant, Cleaner & Deodorizer		According to Manufacturer Spec	01-Feb-18
Pro Restore® Products	Microban X-590 Institutional Spray	Diisobutylphenoxyethoxy ethyl dimethyl benzyl ammonium chloride	According to Manufacturer Spec	11-Aug-16
Anti-Slip Anywhere	Multi Kleen		According to Manufacturer Spec	15-Mar-11
Certo®	Non Acid Bathroom Cleaner		According to Manufacturer Spec	07-May-12
PreVasive®	Noroxycdiff™	Hydrogen Peroxide	According to Manufacturer Spec	22-May-19

[%] Tel: 64 9 274 0869

^{⊚ 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

^{⊕ 1} Spireton Place, Pendle Hill, NSW 2145, Australia

PO Box 335, Pendle Hill, NSW 2145, Australia





Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
Ecolab®	Oasis Pro 12 Neutral Cleaner	Alcohols <10%	Minimum 20 to 1 dilution with water	06-Mar-12
Whiteley Medical	Odacon Spray		According to Manufacturer Spec	14-Jun-11
OdorXit Naturals	Odor Eliminator Spray		Ready to Use	22-Feb-18
Pal® International	PAL® Detergent Wipes		Ready to Use	27-May-09
Pal® International	PAL® Disinfectant Wipes		Ready to Use	27-May-09
PCS Process Cleaning Solutions	PCS 7000	Sodium Hypochlorite	According to Manufacturer Spec	06-Oct-17
PCS Process Cleaning Solutions	PCS Spray & Wipe Hand and Surface Cleaner		Ready to Use	06-Oct-17
Pure Green, LLC	Puregreen 24 (Pure Green)		According to Manufacturer Spec	30-Jan-09
Reynard Health Supplies	Reynard Neutral Detergent Wipes		According to Manufacturer Spec	08-Feb-07
Reynard Health Supplies	Reynard Surface Sanitizing Wipes		According to Manufacturer Spec	27-May-11
Sealed Air Diversey Care	Room Care R2L		According to Manufacturer Spec	24-Oct-12
STERI-7®	S-7 Xtra Wipes		Ready to Use	28-Nov-16
Alco Mandala Products	Sanni-Rinse		According to Manufacturer Spec	16-Feb-12
Clinell®	Sporicidal Wipes		According to Manufacturer Spec	27-May-09
Ecolab®	Spring Clean		Minimum 20 to 1 dilution with water	11-Apr-06
Noble Pine Products Co.	Steri-Fab®	Isopropyl Alchohol	According to Manufacturer Spec	11-Aug-16

[∜] Tel: 64 9 274 0869

^{0 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

^{⊕ 1} Spireton Place, Pendle Hill, NSW 2145, Australia

[☑] PO Box 335, Pendle Hill, NSW 2145, Australia





Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
sBioMed, LLC	Steriplex SD Disinfectant Sporicide Wipes (Without Bleach)	Hydrogen Peroxide	According to Manufacturer Spec	20-May-15
Sterri-Matt [®]	Sterri-Matt® (Pure) Foam		Ready to Use	15-Dec-09
Spartan Chemical Company Inc.	Super HDQ Neutral®		Minimum 256 to 1 dilution with water	11-Aug-16
PDI	Super Sani Cloth® Germicidal Disposable Wipes		Ready to Use	02-Sep-10
Reynard Health Supplies	Surface Sanitising Wipes		Ready to Use	18-Feb-09
3M	TB Quat Disinfectant	Ammonium Chloride	Ready to Use	09-Mar-18
Apollo Corporation	Therasol™		According to Manufacturer Spec	25-Sep-12
Medline Industries Inc.	Touch Screen Cleaning Wipes	70% Isopropyl Alcohol	Ready to Use	15-May-19
(Various)	Trix (Soap)		Ready to Use	27-Apr-10
Verna Care	Tuffie 5 Cleaning Wipes	Cocoaklyl Dimthybenzyl Ammonium Chloride 0-0.5%	Ready to Use	19-Jul-12
Apollo Corporation	Turbo Clean™	Ethanol 9.5%	According to Manufacturer Spec	25-Sep-12
Clinell®	Universal Disinfectant Wipes		According to Manufacturer Spec	27-May-09
Clinell®	Universal Wipes		Ready to Use	12-May-19
True Blue Chemicals	Use All	Glycol 1-6%	According to Manufacturer Spec	28-Sep-12
Whiteley Medical	Viraclean®	Benzalkoniumchloride	According to Manufacturer Spec	19-Jul-12
Sealed Air Diversey Care	Virex® II 256		According to Manufacturer Spec	07-May-12

[%] Tel: 64 9 274 0869

^{⊚ 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

^{⊕ 1} Spireton Place, Pendle Hill, NSW 2145, Australia

PO Box 335, Pendle Hill, NSW 2145, Australia



Approved Cleaners March 2021

PLEASE REFER TO IMPORTANT MESSAGES ON FINAL PAGE BEFORE APPLICATION OF ANY CLEANER/DISINFECTANT.

Sorted by Product Name:

MANUFACTURER	PRODUCT	MAIN ACTIVE COMPONENT	CONCENTRATION	DATE TESTED
Sealed Air Diversey Care	Virex® Plus	Ammonium Chloride	According to Manufacturer Spec	19-Dec-18
Dupont™	Virkon®		Minimum 100 to 1 dilution with water	15-Dec-09
Wexford Labs	Wexcide 128		Minimum 128 to 1 dilution with water	11-Aug-16

^{⊚ 6} Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

Important Messages



CAUTION!

To avoid premature degradation of the surface coating never use an abrasive cloth, sponge or other similar device on the surface. Carefully follow manufacturer instructions for recommended exposure time of cleaner on surface. Wipe off cleaning product residue with a clean, soft cloth after cleaning and allow the surface to dry completely before use.

IMPORTANT INFORMATION:

Cleaners/disinfectants can cause discolouration/bleaching on some colours. To avoid discolouration, always check on a small area first.

Information contained here is purely related to the compatibility of cleaners with **Carrflex**® materials and in no way relates to the efficacy of any cleaner for infection control or the safety of any cleaner for skin contact.

Testing info is specific to the products or formulations listed and in no way constitutes approval for other products from any brand listed without being tested.

All approvals are based on manufacturer recommended dilution with water unless otherwise specified and approval is void for any improper dilution.

Never mix different cleaners. Results can be deadly.

If you wish to change from one approved cleaner to another, thoroughly rinse the surface with warm water, allow to dry completely and repeat this process one more time prior to use of another cleaner.

If a particular cleaner/disinfectant is not listed, we advise to test the specific product in advance or to contact us to conduct testing. Listed chlorine based cleaners should be used only in instances of irregular spot cleaning and bodily fluid spills.

 ⁶ Ra Ora Drive, East Tamaki, Auckland 2013, New Zealand

PO Box 204110, Highbrook, Auckland 2161, New Zealand

PO Box 335, Pendle Hill, NSW 2145, Australia

[△] info@carrgroup.com.au